

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091119\
 Data File : VX012305.D
 Acq On : 11 Sep 2019 16:33
 Operator : SY/SP
 Sample : K4643-14
 Misc : 5.14g/5.0mL/MSVOA X/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T11-C4-(0)

Quant Time: Sep 12 06:37:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 12:55:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	169133	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	288120	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	252113	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	106073	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	113904	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
35) Dibromofluoromethane	5.48	113	87526	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
50) Toluene-d8	8.71	98	344136	53.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.24%	
62) 4-Bromofluorobenzene	11.13	95	126513	42.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.26%	
Target Compounds						
16) Acetone	2.43	43	27908	49.301	ug/l	99
29) Tetrahydrofuran	5.13	42	4244	10.870	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091119\
Data File : VX012305.D
Acq On : 11 Sep 2019 16:33
Operator : SY/SP
Sample : K4643-14
Misc : 5.14g/5.0mL/MSVOA X/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T11-C4-(0)

Quant Time: Sep 12 06:37:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 12:55:26 2019
Response via : Initial Calibration

