

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091119\
 Data File : VX012310.D
 Acq On : 11 Sep 2019 18:54
 Operator : SY/SP
 Sample : K4643-20
 Misc : 2.72g/5.0mL/MSVOA X/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T11-D1-(0)

Quant Time: Sep 12 06:47:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 12:55:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	166944	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	291613	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	248163	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	103301	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	114025	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
35) Dibromofluoromethane	5.48	113	87735	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
50) Toluene-d8	8.71	98	343638	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
62) 4-Bromofluorobenzene	11.14	95	127181	41.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.68%	
Target Compounds						
16) Acetone	2.43	43	19934	35.676	ug/l	94
29) Tetrahydrofuran	5.12	42	4437	11.513	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091119\
Data File : VX012310.D
Acq On : 11 Sep 2019 18:54
Operator : SY/SP
Sample : K4643-20
Misc : 2.72g/5.0mL/MSVOA X/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T11-D1-(0)

Quant Time: Sep 12 06:47:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 12:55:26 2019
Response via : Initial Calibration

