

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091119\
 Data File : VX012319.D
 Acq On : 11 Sep 2019 23:57
 Operator : SY/SP
 Sample : K4789-12
 Misc : 5.55g/5.0mL/MSVOA X/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RT3860

Quant Time: Sep 12 07:02:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 12:55:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	116787	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	189102	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	164061	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	74393	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	71284	43.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.34%	
35) Dibromofluoromethane	5.48	113	52504	43.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.62%	
50) Toluene-d8	8.71	98	223328	52.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.06%	
62) 4-Bromofluorobenzene	11.14	95	85501	43.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.76%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091119\
Data File : VX012319.D
Acq On : 11 Sep 2019 23:57
Operator : SY/SP
Sample : K4789-12
Misc : 5.55g/5.0mL/MSVOA X/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RT3860

Quant Time: Sep 12 07:02:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 12:55:26 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 781

Delta R.T. -0.00 min

Lab File: VX012319.D

Acq: 11 Sep 2019 23:57

Instrument :

MSVOA_X

ClientSampleId :


RT3860

Tot Ion:168 Resp: 116787

Ion Ratio Lower Upper

168 100

99 65.9 55.8 83.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65

40000

30000



20000

10000

0

Time-->

5.60 5.70



#33

1,2-Dichloroethane-d4

Concen: 43.173 ug/l

RT: 6.05 min Scan# 847

Delta R.T. -0.00 min

Lab File: VX012319.D


Acq: 11 Sep 2019 23:57

Tgt Ion: 65 Resp: 71284

Ion Ratio Lower Upper

65 100

67 49.0 0.0 99.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.05

30000

25000

20000

15000


10000


5000

0

Time-->

6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. -0.00 min

Lab File: VX012319.D

Acq: 11 Sep 2019 23:57

Instrument :

MSVOA_X

ClientSampled :

RT3860


Tot Ion:114 Resp: 189102

Ion Ratio Lower Upper

114 100

63 22.1 0.0 45.2




88 16.7 0.0 34.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 43.312 ug/l

RT: 5.48 min Scan# 754

Delta R.T. -0.00 min

Lab File: VX012319.D

Acq: 11 Sep 2019 23:57


Tgt Ion:113 Resp: 52504

Ion Ratio Lower Upper

113 100

111 105.1 83.0 124.6



192 18.4 14.3 21.5




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 52.527 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. -0.00 min
Lab File: VX012319.D
Acq: 11 Sep 2019 23:57




Instrument :
MSVOA_X
ClientSampleId :
RT3860


Tot Ion: 98 Resp: 22328
Ion Ratio Lower Upper
98 100
100 66.5 54.1 81.1



#62
4-Bromofluorobenzene
Concen: 43.377 ug/l
RT: 11.14 min Scan# 1681
Delta R.T. 0.01 min
Lab File: VX012319.D
Acq: 11 Sep 2019 23:57

Tgt Ion: 95 Resp: 85501
Ion Ratio Lower Upper
95 100
174 66.7 0.0 127.0
176 64.2 0.0 125.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. -0.00 min
Lab File: VX012319.D
Acq: 11 Sep 2019 23:57

Instrument :
MSVOA_X
ClientSampled :
RT3860

Tot Ion	Ratio	Lower	Upper
117	100		
82	56.6	46.4	69.6
119	31.2	26.6	39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. -0.00 min
Lab File: VX012319.D
Acq: 11 Sep 2019 23:57

Tgt Ion	Ratio	Lower	Upper
152	100		
115	64.8	45.5	136.4
150	156.5	0.0	346.4

