

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091119\
 Data File : VX012330.D
 Acq On : 12 Sep 2019 05:06
 Operator : SY/SP
 Sample : K4694-14
 Misc : 3.73g/5.0mL/MSVOA X/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T2-2-(1.5)

Quant Time: Sep 12 07:26:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 12:55:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	156735	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	269331	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	189399	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	50425	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	107399	48.47	ug/l	0.00
Spiked Amount						
			Recovery	=	96.94%	
35) Dibromofluoromethane	5.48	113	80508	46.63	ug/l	0.00
Spiked Amount						
			Recovery	=	93.26%	
50) Toluene-d8	8.71	98	295514	48.80	ug/l	0.00
Spiked Amount						
			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	11.14	95	80009	28.50	ug/l	0.00
Spiked Amount						
			Recovery	=	57.00%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091119\
Data File : VX012330.D
Acq On : 12 Sep 2019 05:06
Operator : SY/SP
Sample : K4694-14
Misc : 3.73g/5.0mL/MSVOA X/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T2-2-(1.5)

Quant Time: Sep 12 07:26:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 12:55:26 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 781

Delta R.T. 0.00 min

Lab File: VX012330.D

Acq: 12 Sep 2019 05:06

Instrument :

MSVOA_X

ClientSampleId :


T2-2-(1.5)

Tot Ion:168 Resp: 156735

Ion Ratio Lower Upper

168 100

99 68.0 55.8 83.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65

60000

50000

40000



30000

20000

10000

0

Time--> 5.50 5.60 5.70



#33

1,2-Dichloroethane-d4

Concen: 48.467 ug/l

RT: 6.05 min Scan# 847

Delta R.T. 0.00 min

Lab File: VX012330.D


Acq: 12 Sep 2019 05:06

Tgt Ion: 65 Resp: 107399

Ion Ratio Lower Upper

65 100

67 49.2 0.0 99.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.05

40000


30000


20000

10000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. 0.00 min

Lab File: VX012330.D

Acq: 12 Sep 2019 05:06

Instrument :

MSVOA_X

ClientSampleId :

T2-2-(1.5)


Tot Ion:114 Resp: 269331

Ion Ratio Lower Upper

114 100

63 22.2 0.0 45.2




88 17.5 0.0 34.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 46.630 ug/l

RT: 5.48 min Scan# 754

Delta R.T. 0.00 min

Lab File: VX012330.D

Acq: 12 Sep 2019 05:06


Tgt Ion:113 Resp: 80508

Ion Ratio Lower Upper

113 100

111 101.2 83.0 124.6



192 17.5 14.3 21.5




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.800 ug/l

RT: 8.71 min Scan# 1283

Delta R.T. 0.00 min

Lab File: VX012330.D

Acq: 12 Sep 2019 05:06

Instrument :

MSVOA_X

ClientSampleId :


T2-2-(1.5)

Tot Ion: 98 Resp: 295514

Ion Ratio Lower Upper

98 100

100 66.4 54.1 81.1



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX

8.71

200000

150000



100000

50000

0

Time-->

8.60 8.70 8.80 8.90



#62

4-Bromofluorobenzene

Concen: 28.500 ug/l

RT: 11.14 min Scan# 1681

Delta R.T. 0.01 min

Lab File: VX012330.D

Acq: 12 Sep 2019 05:06


Tgt Ion: 95 Resp: 80009

Ion Ratio Lower Upper

95 100

174 67.2 0.0 127.0

176 65.1 0.0 125.8



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

11.14

80000

60000


40000


20000

0

Time-->

11.10 11.20









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. 0.00 min
Lab File: VX012330.D
Acq: 12 Sep 2019 05:06

Instrument :
MSVOA_X
ClientSampleId :
T2-2-(1.5)

Tot Ion	Ratio	Lower	Upper
117	100		
82	57.9	46.4	69.6
119	32.4	26.6	39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. 0.00 min
Lab File: VX012330.D
Acq: 12 Sep 2019 05:06

Tgt Ion	Ratio	Lower	Upper
152	100		
115	65.1	45.5	136.4
150	157.5	0.0	346.4

