

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091120\
 Data File : VX018367.D
 Acq On : 11 Sep 2020 17:44
 Operator : JC/SP
 Sample : PB131535TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131535TB

Quant Time: Sep 12 07:13:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	441012	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	705130	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	693236	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	330241	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	312878	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
35) Dibromofluoromethane	5.46	113	247333	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
50) Toluene-d8	8.70	98	879436	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
62) 4-Bromofluorobenzene	11.12	95	326503	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
Target Compounds						
16) Acetone	2.42	43	21337	8.944	ug/l	97
20) Methylene Chloride	2.84	84	14632	1.876	ug/l #	92
43) Isopropyl Acetate	6.42	43	117957	9.668	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091120\
Data File : VX018367.D
Acq On : 11 Sep 2020 17:44
Operator : JC/SP
Sample : PB131535TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131535TB

Quant Time: Sep 12 07:13:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

