

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091218\
 Data File : VX004493.D
 Acq On : 12 Sep 2018 14:34
 Operator : JC/MD
 Sample : J4848-03 10X
 Misc : 5.05µ/5mL/100µL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-MW-A(12-16)

Quant Time: Sep 13 11:48:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	196384	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	310035	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	322243	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	198844	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	135929	53.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.36%	
35) Dibromofluoromethane	5.50	113	117494	41.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.82%	
50) Toluene-d8	8.72	98	463078	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
62) 4-Bromofluorobenzene	11.14	95	244435	71.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	143.94%	
Target Compounds						
73) Isopropylbenzene	11.02	105	334152	25.52	ug/l	99
78) n-propylbenzene	11.37	91	852568	56.62	ug/l	98
85) sec-Butylbenzene	11.95	105	658546	49.31	ug/l	94
89) n-Butylbenzene	12.39	91	370230	34.45	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091218\
Data File : VX004493.D
Acq On : 12 Sep 2018 14:34
Operator : JC/MD
Sample : J4848-03 10X
Misc : 5.05µ/5mL/100µL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-MW-A(12-16)

Quant Time: Sep 13 11:48:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

