

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091218\
 Data File : VX004498.D
 Acq On : 12 Sep 2018 16:49
 Operator : JC/MD
 Sample : J4724-16 50X
 Misc : 4.95g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-SUT-4USTS-S

Quant Time: Sep 13 11:59:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	208135	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	330391	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	341123	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	196391	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	144426	53.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.64%	
35) Dibromofluoromethane	5.50	113	125757	41.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.18%	
50) Toluene-d8	8.72	98	482433	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
62) 4-Bromofluorobenzene	11.14	95	198443	54.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.66%	
Target Compounds						
73) Isopropylbenzene	11.02	105	35809	2.77	ug/l	Qvalue 100
80) 1,3,5-Trimethylbenzene	11.51	105	386326	35.53	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	1159327	104.43	ug/l	100
85) sec-Butylbenzene	11.95	105	120015	9.10	ug/l #	78
86) p-Isopropyltoluene	12.07	119	154335	13.25	ug/l	100
89) n-Butylbenzene	12.38	91	215317	20.28	ug/l #	30
95) Naphthalene	13.83	128	61628	4.60	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091218\
Data File : VX004498.D
Acq On : 12 Sep 2018 16:49
Operator : JC/MD
Sample : J4724-16 50X
Misc : 4.95g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-SUT-4USTS-S

Quant Time: Sep 13 11:59:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

