

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091218\
 Data File : VX004501.D
 Acq On : 12 Sep 2018 18:09
 Operator : JC/MD
 Sample : J4848-01DL 100X
 Misc : 5.67g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-MW-A(6-8)DL

Quant Time: Sep 13 11:46:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	211293	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	334794	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	338435	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	190220	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	145459	53.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.78%	
35) Dibromofluoromethane	5.51	113	127981	41.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.54%	
50) Toluene-d8	8.72	98	491347	47.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.64%	
62) 4-Bromofluorobenzene	11.14	95	194024	52.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.82%	
Target Compounds						
73) Isopropylbenzene	11.02	105	78060	6.23	ug/l	Qvalue 100
78) n-propylbenzene	11.36	91	220440	15.30	ug/l	98
84) 1,2,4-Trimethylbenzene	11.81	105	739439	68.77	ug/l	100
85) sec-Butylbenzene	11.94	105	131610	10.30	ug/l	80
86) p-Isopropyltoluene	12.07	119	43002	3.81	ug/l	99
89) n-Butylbenzene	12.39	91	140594	13.67	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091218\
Data File : VX004501.D
Acq On : 12 Sep 2018 18:09
Operator : JC/MD
Sample : J4848-01DL 100X
Misc : 5.67µ/5mL/100µL/5.00mL/MSVOA_X/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-MW-A(6-8)DL

Quant Time: Sep 13 11:46:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

