

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091218\
 Data File : VX004504.D
 Acq On : 12 Sep 2018 19:30
 Operator : JC/MD
 Sample : J4724-18DL 100X
 Misc : 3.19g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-SUT-4USTS-SBDL

Quant Time: Sep 13 03:25:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	218288	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	344087	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	343827	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	198224	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	147975	52.57	ug/l	0.00
Spiked Amount			Recovery	=	105.14%	
35) Dibromofluoromethane	5.51	113	130283	41.38	ug/l	0.00
Spiked Amount			Recovery	=	82.76%	
50) Toluene-d8	8.72	98	492799	46.67	ug/l	0.00
Spiked Amount			Recovery	=	93.34%	
62) 4-Bromofluorobenzene	11.14	95	195432	51.85	ug/l	0.00
Spiked Amount			Recovery	=	103.70%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
67) Ethyl Benzene	10.26	91	23016	1.66	ug/l	99
68) m/p-Xylenes	10.36	106	12090	2.26	ug/l	94
73) Isopropylbenzene	11.02	105	79072	6.06	ug/l	100
78) n-propylbenzene	11.36	91	231382	15.41	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	445558	40.60	ug/l	97
83) tert-Butylbenzene	11.77	119	19055	1.82	ug/l	88
84) 1,2,4-Trimethylbenzene	11.81	105	1449238	129.34	ug/l	100
85) sec-Butylbenzene	11.95	105	139397	10.47	ug/l	79
86) p-Isopropyltoluene	12.07	119	110124	9.37	ug/l	99
89) n-Butylbenzene	12.39	91	188257	17.57	ug/l #	46
95) Naphthalene	13.83	128	82081	6.07	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091218\
Data File : VX004504.D
Acq On : 12 Sep 2018 19:30
Operator : JC/MD
Sample : J4724-18DL 100X
Misc : 3.19g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-SUT-4USTS-SBDL

Quant Time: Sep 13 03:25:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

