

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
 Data File : VX031274.D
 Acq On : 12 Sep 2022 13:08
 Operator : JC/MD
 Sample : VX0912WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0912WBL01

Quant Time: Sep 13 01:33:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	72881	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	114100	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86214	49.965	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.920%
35) Dibromofluoromethane	5.385	113	60382	48.563	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.120%
50) Toluene-d8	8.646	98	201293	44.559	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.120%
62) 4-Bromofluorobenzene	11.079	95	76702	46.331	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.660%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
Data File : VX031274.D
Acq On : 12 Sep 2022 13:08
Operator : JC/MD
Sample : VX0912WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0912WBL01

Quant Time: Sep 13 01:33:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

