

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
 Data File : VX031277.D
 Acq On : 12 Sep 2022 15:42
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 13 01:35:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	78296	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	6.757	114	125365	50.000	ug/l	-0.01
63) Chlorobenzene-d5	10.055	117	113665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58444	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	101906	54.974	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.940%
35) Dibromofluoromethane	5.385	113	68342	50.026	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.060%
50) Toluene-d8	8.647	98	226620	45.657	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.320%
62) 4-Bromofluorobenzene	11.079	95	90680	49.853	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
Data File : VX031277.D
Acq On : 12 Sep 2022 15:42
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 13 01:35:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

