

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
 Data File : VX031278.D
 Acq On : 12 Sep 2022 16:05
 Operator : JC/MD
 Sample : N4528-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PIT-10

Quant Time: Sep 13 01:36:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	73393	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	115088	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	103177	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86595	49.836	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.680%
35) Dibromofluoromethane	5.385	113	60564	48.291	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	8.653	98	199478	43.778	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.560%
62) 4-Bromofluorobenzene	11.079	95	77121	46.184	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.360%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	13621	22.639	ug/l	# 87
20) Methylene Chloride	2.788	84	3858	2.728	ug/l	# 85
37) Ethyl Acetate	4.721	43	6344	3.807	ug/l	# 94
43) Isopropyl Acetate	6.342	43	29229	11.126	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
Data File : VX031278.D
Acq On : 12 Sep 2022 16:05
Operator : JC/MD
Sample : N4528-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PIT-10

Quant Time: Sep 13 01:36:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

