

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
 Data File : VX031282.D
 Acq On : 12 Sep 2022 18:36
 Operator : JC/MD
 Sample : N4536-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 090722-C

Quant Time: Sep 13 01:57:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	69358	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	109507	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	98118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45371	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	84910	51.709	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.420%	
35) Dibromofluoromethane	5.397	113	57021	47.783	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.560%	
50) Toluene-d8	8.653	98	191706	44.216	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 88.440%	
62) 4-Bromofluorobenzene	11.085	95	74135	46.659	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.320%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	14493	25.490	ug/l	# 88
20) Methylene Chloride	2.794	84	3519	2.633	ug/l	89
37) Ethyl Acetate	4.727	43	10007	6.312	ug/l	# 92
40) Benzene	6.044	78	65879	13.570	ug/l	99
43) Isopropyl Acetate	6.354	43	40914	16.368	ug/l	# 93
52) Toluene	8.726	92	59603	18.614	ug/l	96
67) Ethyl Benzene	10.195	91	12526	2.393	ug/l	# 88
68) m/p-Xylenes	10.305	106	24959	12.861	ug/l	92
69) o-Xylene	10.646	106	18464	9.533	ug/l	86
70) Styrene	10.659	104	8158	2.627	ug/l	# 62
80) 1,3,5-Trimethylbenzene	11.457	105	6680	2.017	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	20229	6.282	ug/l	84
95) Naphthalene	13.780	128	255052	71.340	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
Data File : VX031282.D
Acq On : 12 Sep 2022 18:36
Operator : JC/MD
Sample : N4536-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
090722-C

Quant Time: Sep 13 01:57:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

