

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
 Data File : VX031285.D
 Acq On : 12 Sep 2022 19:47
 Operator : JC/MD
 Sample : N4298-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20220822

Quant Time: Sep 13 01:59:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	69469	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	111109	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	98074	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	80634	49.026	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.060%
35) Dibromofluoromethane	5.391	113	55338	45.704	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.400%
50) Toluene-d8	8.653	98	126668	28.794	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	57.580%#
62) 4-Bromofluorobenzene	11.085	95	63408	39.332	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	78.660%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
Data File : VX031285.D
Acq On : 12 Sep 2022 19:47
Operator : JC/MD
Sample : N4298-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20220822

Quant Time: Sep 13 01:59:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

