

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
 Data File : VX031286.D
 Acq On : 12 Sep 2022 20:11
 Operator : JC/MD
 Sample : N4298-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW181S-20220822

Quant Time: Sep 13 01:59:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	70019	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	111767	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	98881	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44046	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	84689	51.087	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.180%
35) Dibromofluoromethane	5.391	113	57843	47.492	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.980%
50) Toluene-d8	8.653	98	189742	42.878	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	85.760%#
62) 4-Bromofluorobenzene	11.085	95	72589	44.762	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.520%
Target Compounds						
						Qvalue
44) Trichloroethene	7.123	130	926	0.665	ug/l	50
64) Tetrachloroethene	9.275	164	1909	1.855	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091222\
Data File : VX031286.D
Acq On : 12 Sep 2022 20:11
Operator : JC/MD
Sample : N4298-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW181S-20220822

Quant Time: Sep 13 01:59:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

