

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
 Data File : VX037657.D
 Acq On : 12 Sep 2023 10:22
 Operator : JC/MD
 Sample : VX0912MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0912MBL01

Quant Time: Sep 12 10:47:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	183922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	319099	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	287402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120723	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	126595	42.543	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.080%
35) Dibromofluoromethane	5.385	113	100464	44.418	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.840%
50) Toluene-d8	8.647	98	391253	45.867	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.740%#
62) 4-Bromofluorobenzene	11.079	95	132404	43.023	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
Data File : VX037657.D
Acq On : 12 Sep 2023 10:22
Operator : JC/MD
Sample : VX0912MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0912MBL01

Quant Time: Sep 12 10:47:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

