

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
 Data File : VX037661.D
 Acq On : 12 Sep 2023 12:09
 Operator : JC/MD
 Sample : 04327-01 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-10A-9.5

Quant Time: Sep 13 01:43:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	170550	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	284811	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	259739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	113911	41.281	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.560%
35) Dibromofluoromethane	5.385	113	88788	43.981	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.960%
50) Toluene-d8	8.647	98	351373	46.150	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.300%
62) 4-Bromofluorobenzene	11.079	95	117132	42.643	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
Data File : VX037661.D
Acq On : 12 Sep 2023 12:09
Operator : JC/MD
Sample : 04327-01 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-10A-9.5

Quant Time: Sep 13 01:43:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

