

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
 Data File : VX037665.D
 Acq On : 12 Sep 2023 13:46
 Operator : JC/MD
 Sample : 04297-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1731-DRIP

Quant Time: Sep 12 14:26:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	187418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	322557	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	295374	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	127283	41.976	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.960%
35) Dibromofluoromethane	5.385	113	100808	44.092	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.180%
50) Toluene-d8	8.647	98	394608	45.764	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.520%#
62) 4-Bromofluorobenzene	11.079	95	133976	43.067	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.140%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	30794	24.566	ug/l	92
17) Carbon Disulfide	2.508	76	7032	1.431	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
Data File : VX037665.D
Acq On : 12 Sep 2023 13:46
Operator : JC/MD
Sample : 04297-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1731-DRIP

Quant Time: Sep 12 14:26:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

