

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
 Data File : VX037670.D
 Acq On : 12 Sep 2023 15:48
 Operator : JC/MD
 Sample : 04373-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1044-MW-11(15)

Quant Time: Sep 13 01:46:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	210138	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	370402	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	344335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138548	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	155902	45.855	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	91.720%	
35) Dibromofluoromethane	5.385	113	121255	46.184	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	92.360%	
50) Toluene-d8	8.647	98	460107	46.467	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	92.940%	
62) 4-Bromofluorobenzene	11.079	95	152949	42.816	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	85.640%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	12336	8.975	ug/l	90
25) 2-Butanone	4.574	43	14201	6.336	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	2294	0.756	ug/l	91
44) Trichloroethene	7.129	130	5346	1.976	ug/l	96
51) 4-Methyl-2-Pentanone	8.580	43	19956	4.661	ug/l	98
64) Tetrachloroethene	9.275	164	24078	10.793	ug/l	98
95) Naphthalene	13.780	128	2540	0.249	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
Data File : VX037670.D
Acq On : 12 Sep 2023 15:48
Operator : JC/MD
Sample : 04373-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1044-MW-11(15)

Quant Time: Sep 13 01:46:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

