

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
 Data File : VX037671.D
 Acq On : 12 Sep 2023 16:13
 Operator : JC/MD
 Sample : 04256-03ME
 Misc : 6.36g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2ME

Quant Time: Sep 13 01:46:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	158519	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	265255	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97544	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103550	40.375	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	80.740%
35) Dibromofluoromethane	5.391	113	78383	41.690	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	83.380%
50) Toluene-d8	8.647	98	319880	45.111	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.220%#
62) 4-Bromofluorobenzene	11.085	95	106876	41.778	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.560%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.776	84	4533	1.013	ug/l #	88
39) Methylcyclohexane	7.379	83	4032	1.332	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
Data File : VX037671.D
Acq On : 12 Sep 2023 16:13
Operator : JC/MD
Sample : 04256-03ME
Misc : 6.36g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2ME

Quant Time: Sep 13 01:46:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

