

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
 Data File : VX037676.D
 Acq On : 12 Sep 2023 18:15
 Operator : JC/MD
 Sample : 04397-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Sep 13 01:48:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	208905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	373719	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	345212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139792	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156310	46.246	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.500%
35) Dibromofluoromethane	5.385	113	121119	45.723	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.440%
50) Toluene-d8	8.646	98	460953	46.140	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.280%
62) 4-Bromofluorobenzene	11.079	95	151084	41.918	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091223\
Data File : VX037676.D
Acq On : 12 Sep 2023 18:15
Operator : JC/MD
Sample : 04397-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Sep 13 01:48:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

