

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091319\
 Data File : VX012383.D
 Acq On : 13 Sep 2019 13:34
 Operator : SY/SP
 Sample : K4695-03
 Misc : 4.79g/5.0mL/MSVOA X/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-2-(1.5)

Quant Time: Sep 14 00:19:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 12:55:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	165642	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	280672	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	233179	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	84917	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	106418	45.44	ug/l	0.00
Spiked Amount			Recovery	=	90.88%	
35) Dibromofluoromethane	5.48	113	80777	44.90	ug/l	0.00
Spiked Amount			Recovery	=	89.80%	
50) Toluene-d8	8.71	98	327502	51.90	ug/l	0.00
Spiked Amount			Recovery	=	103.80%	
62) 4-Bromofluorobenzene	11.14	95	112483	38.45	ug/l	0.00
Spiked Amount			Recovery	=	76.90%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091319\
Data File : VX012383.D
Acq On : 13 Sep 2019 13:34
Operator : SY/SP
Sample : K4695-03
Misc : 4.79g/5.0mL/MSVOA X/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-2-(1.5)

Quant Time: Sep 14 00:19:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 12:55:26 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 781

Delta R.T. -0.00 min

Lab File: VX012383.D

Acq: 13 Sep 2019 13:34

Instrument :

MSVOA_X

ClientSampleId :


T3-2-(1.5)

Tot Ion:168 Resp: 165642

Ion Ratio Lower Upper




168 100

99 67.2 55.8 83.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 45.442 ug/l

RT: 6.05 min Scan# 847

Delta R.T. 0.00 min

Lab File: VX012383.D


Acq: 13 Sep 2019 13:34

Tgt Ion: 65 Resp: 106418

Ion Ratio Lower Upper



65 100


67 50.2 0.0 99.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. 0.00 min

Lab File: VX012383.D

Acq: 13 Sep 2019 13:34

Instrument :

MSVOA_X

ClientSampleId :

T3-2-(1.5)


Tot Ion:114 Resp: 280672

Ion Ratio Lower Upper

114 100

63 21.4 0.0 45.2




88 17.2 0.0 34.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 44.895 ug/l

RT: 5.48 min Scan# 754

Delta R.T. -0.00 min

Lab File: VX012383.D

Acq: 13 Sep 2019 13:34


Tgt Ion:113 Resp: 80777

Ion Ratio Lower Upper

113 100

111 104.1 83.0 124.6



192 18.3 14.3 21.5




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.898 ug/l

RT: 8.71 min Scan# 1283

Delta R.T. -0.00 min

Lab File: VX012383.D

Acq: 13 Sep 2019 13:34

Instrument :

MSVOA_X

ClientSampleId :


T3-2-(1.5)

Tot Ion: 98 Resp: 327502

Ion Ratio Lower Upper


98 100

100 66.2 54.1 81.1





Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



Time-->



#62

4-Bromofluorobenzene

Concen: 38.448 ug/l

RT: 11.14 min Scan# 1681

Delta R.T. 0.01 min

Lab File: VX012383.D

Acq: 13 Sep 2019 13:34


Tgt Ion: 95 Resp: 112483

Ion Ratio Lower Upper

95 100

174 67.9 0.0 127.0


176 65.8 0.0 125.8




Abundance Ion 94.90 (94.60 to 95.60): VX0


Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V



Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. -0.00 min
Lab File: VX012383.D
Acq: 13 Sep 2019 13:34

Instrument :
MSVOA_X
ClientSampleId :
T3-2-(1.5)

Tot Ion:117 Resp: 233179
Ion Ratio Lower Upper
117 100
82 56.2 46.4 69.6
119 32.7 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. -0.00 min
Lab File: VX012383.D
Acq: 13 Sep 2019 13:34

Tgt Ion:152 Resp: 84917
Ion Ratio Lower Upper
152 100
115 62.7 45.5 136.4
150 153.5 0.0 346.4

