

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091319\
 Data File : VX012389.D
 Acq On : 13 Sep 2019 16:21
 Operator : SY/SP
 Sample : K4695-10
 Misc : 5.10g/5.0mL/MSVOA X/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T5-2-(2.5)

Quant Time: Sep 14 00:38:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 12:55:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	161721	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	275006	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	240457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	103736	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	106466	46.56	ug/l	0.00
Spiked Amount			Recovery	=	93.12%	
35) Dibromofluoromethane	5.48	113	80601	45.72	ug/l	0.00
Spiked Amount			Recovery	=	91.44%	
50) Toluene-d8	8.71	98	324710	52.52	ug/l	0.00
Spiked Amount			Recovery	=	105.04%	
62) 4-Bromofluorobenzene	11.14	95	120770	42.13	ug/l	0.00
Spiked Amount			Recovery	=	84.26%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091319\
Data File : VX012389.D
Acq On : 13 Sep 2019 16:21
Operator : SY/SP
Sample : K4695-10
Misc : 5.10g/5.0mL/MSVOA X/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T5-2-(2.5)

Quant Time: Sep 14 00:38:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 12:55:26 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 781

Delta R.T. 0.00 min

Lab File: VX012389.D

Acq: 13 Sep 2019 16:21

Instrument :

MSVOA_X

ClientSampleId :


T5-2-(2.5)

Tot Ion:168 Resp: 161721

Ion Ratio Lower Upper




168 100

99 65.9 55.8 83.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 46.565 ug/l

RT: 6.05 min Scan# 847

Delta R.T. 0.00 min

Lab File: VX012389.D


Acq: 13 Sep 2019 16:21

Tgt Ion: 65 Resp: 106466

Ion Ratio Lower Upper



65 100


67 51.0 0.0 99.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. 0.00 min

Lab File: VX012389.D

Acq: 13 Sep 2019 16:21

Instrument :

MSVOA_X

ClientSampleId :

T5-2-(2.5)


Tot Ion:114 Resp: 275006

Ion Ratio Lower Upper

114 100

63 22.2 0.0 45.2




88 17.7 0.0 34.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 45.720 ug/l

RT: 5.48 min Scan# 754

Delta R.T. 0.00 min

Lab File: VX012389.D

Acq: 13 Sep 2019 16:21


Tgt Ion:113 Resp: 80601

Ion Ratio Lower Upper

113 100

111 102.4 83.0 124.6



192 18.8 14.3 21.5




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 52.515 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. 0.00 min
Lab File: VX012389.D
Acq: 13 Sep 2019 16:21




Instrument :
MSVOA_X
ClientSampleId :
T5-2-(2.5)


Tot Ion: 98 Resp: 324710
Ion Ratio Lower Upper
98 100
100 66.1 54.1 81.1



#62
4-Bromofluorobenzene
Concen: 42.131 ug/l
RT: 11.14 min Scan# 1681
Delta R.T. 0.01 min
Lab File: VX012389.D
Acq: 13 Sep 2019 16:21

Tgt Ion: 95 Resp: 120770
Ion Ratio Lower Upper
95 100
174 69.3 0.0 127.0
176 66.0 0.0 125.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. 0.00 min
Lab File: VX012389.D
Acq: 13 Sep 2019 16:21

Instrument :
MSVOA_X
ClientSampleId :
T5-2-(2.5)

Tot Ion	Ratio	Lower	Upper
117	100		
82	56.4	46.4	69.6
119	31.5	26.6	39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. 0.00 min
Lab File: VX012389.D
Acq: 13 Sep 2019 16:21

Tgt Ion	Ratio	Lower	Upper
152	100		
115	63.9	45.5	136.4
150	159.5	0.0	346.4

