

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091319\
 Data File : VX012395.D
 Acq On : 13 Sep 2019 19:10
 Operator : SY/SP
 Sample : K4695-04
 Misc : 6.02g/5.0mL/MSVOA X/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-3-(0)

Quant Time: Sep 14 00:49:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 12:55:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	169666	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	287897	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	248673	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	102457	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	110157	45.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.84%	
35) Dibromofluoromethane	5.48	113	84159	45.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.20%	
50) Toluene-d8	8.71	98	335311	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
62) 4-Bromofluorobenzene	11.13	95	121401	40.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.92%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091319\
Data File : VX012395.D
Acq On : 13 Sep 2019 19:10
Operator : SY/SP
Sample : K4695-04
Misc : 6.02g/5.0mL/MSVOA X/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-3-(0)

Quant Time: Sep 14 00:49:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 12:55:26 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 781

Delta R.T. 0.00 min

Lab File: VX012395.D

Acq: 13 Sep 2019 19:10

Instrument :

MSVOA_X

ClientSampleId :


T3-3-(0)

Tot Ion:168 Resp: 169666

Ion Ratio Lower Upper




168 100

99 66.2 55.8 83.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 45.923 ug/l

RT: 6.05 min Scan# 847

Delta R.T. 0.00 min

Lab File: VX012395.D


Acq: 13 Sep 2019 19:10

Tgt Ion: 65 Resp: 110157

Ion Ratio Lower Upper



65 100


67 50.6 0.0 99.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. 0.00 min

Lab File: VX012395.D

Acq: 13 Sep 2019 19:10

Instrument :

MSVOA_X

ClientSampleId :

T3-3-(0)


Tot Ion:114 Resp: 287897

Ion Ratio Lower Upper

114 100

63 21.2 0.0 45.2




88 17.0 0.0 34.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 45.601 ug/l

RT: 5.48 min Scan# 754

Delta R.T. 0.00 min

Lab File: VX012395.D

Acq: 13 Sep 2019 19:10


Tgt Ion:113 Resp: 84159

Ion Ratio Lower Upper

113 100

111 104.3 83.0 124.6



192 18.2 14.3 21.5




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 51.802 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. 0.00 min
Lab File: VX012395.D
Acq: 13 Sep 2019 19:10




Instrument :
MSVOA_X
ClientSampleId :
T3-3-(0)


Tot Ion: 98 Resp: 335311
Ion Ratio Lower Upper
98 100
100 67.7 54.1 81.1



#62
4-Bromofluorobenzene
Concen: 40.455 ug/l
RT: 11.13 min Scan# 1680
Delta R.T. 0.00 min
Lab File: VX012395.D
Acq: 13 Sep 2019 19:10

Tgt Ion: 95 Resp: 121401
Ion Ratio Lower Upper
95 100
174 70.5 0.0 127.0
176 69.1 0.0 125.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. 0.00 min
Lab File: VX012395.D
Acq: 13 Sep 2019 19:10

Instrument :
MSVOA_X
ClientSampleId :
T3-3-(0)

Tot Ion	Ratio	Lower	Upper
117	100		
82	55.9	46.4	69.6
119	32.3	26.6	39.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. 0.00 min
Lab File: VX012395.D
Acq: 13 Sep 2019 19:10

Tgt Ion	Ratio	Lower	Upper
152	100		
115	62.2	45.5	136.4
150	157.4	0.0	346.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

