

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091321\
 Data File : VX024185.D
 Acq On : 13 Sep 2021 15:32
 Operator : JC/MD
 Sample : VX0913WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0913WBL01

Quant Time: Sep 14 05:22:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091321W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 13 12:27:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	153586	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	255227	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83271	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	108444	49.979	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.960%
35) Dibromofluoromethane	5.397	113	85667	49.056	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.120%
50) Toluene-d8	8.653	98	312598	49.285	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.580%
62) 4-Bromofluorobenzene	11.085	95	103800	42.441	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.880%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	1271	1.624	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091321\
Data File : VX024185.D
Acq On : 13 Sep 2021 15:32
Operator : JC/MD
Sample : VX0913WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0913WBL01

Quant Time: Sep 14 05:22:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091321W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 13 12:27:19 2021
Response via : Initial Calibration

