

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091321\
 Data File : VX024190.D
 Acq On : 13 Sep 2021 17:35
 Operator : JC/MD
 Sample : PB139034ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139034ZHE#02

Quant Time: Sep 14 05:25:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091321W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 13 12:27:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	147848	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	255710	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247912	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	108308	51.853	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.700%
35) Dibromofluoromethane	5.397	113	83526	47.739	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.480%
50) Toluene-d8	8.653	98	318862	50.178	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.360%
62) 4-Bromofluorobenzene	11.085	95	118561	48.385	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.760%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	49839	53.626	ug/l	90
20) Methylene Chloride	2.794	84	15078	10.424	ug/l	93
25) 2-Butanone	4.580	43	8979	6.860	ug/l	96
43) Isopropyl Acetate	6.355	43	21296	4.887	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091321\
Data File : VX024190.D
Acq On : 13 Sep 2021 17:35
Operator : JC/MD
Sample : PB139034ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139034ZHE#02

Quant Time: Sep 14 05:25:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091321W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 13 12:27:19 2021
Response via : Initial Calibration

