

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091321\
 Data File : VX024191.D
 Acq On : 13 Sep 2021 17:59
 Operator : JC/MD
 Sample : M3686-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MAIN-GUARD-HOUSE-JET-GROUT

Quant Time: Sep 14 05:25:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091321W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 13 12:27:19 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	238773	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97393	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104038	51.967	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	103.940%	
35) Dibromofluoromethane	5.397	113	33170	20.303	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	40.600%#	
50) Toluene-d8	8.653	98	296554	49.978	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	99.960%	
62) 4-Bromofluorobenzene	11.085	95	110047	48.096	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	96.200%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	200760	256.594	ug/l	90
20) Methylene Chloride	2.794	84	14436	10.414	ug/l	88
25) 2-Butanone	4.574	43	13377	10.663	ug/l	90
95) Naphthalene	13.780	128	10037	1.629	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091321\
Data File : VX024191.D
Acq On : 13 Sep 2021 17:59
Operator : JC/MD
Sample : M3686-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MAIN-GUARD-HOUSE-JET-GROUT

Quant Time: Sep 14 05:25:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091321W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 13 12:27:19 2021
Response via : Initial Calibration

