

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091321\
 Data File : VX024192.D
 Acq On : 13 Sep 2021 18:22
 Operator : JC/MD
 Sample : M3709-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 250EM-DISPOSAL-1

Quant Time: Sep 14 05:25:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091321W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 13 12:27:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136923	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234374	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101630	52.539	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.080%
35) Dibromofluoromethane	5.397	113	77979	48.626	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.260%
50) Toluene-d8	8.653	98	291554	50.057	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.120%
62) 4-Bromofluorobenzene	11.085	95	107830	48.011	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.020%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	27434	27.919	ug/l	92
20) Methylene Chloride	2.794	84	8178	6.456	ug/l	96
27) cis-1,2-Dichloroethene	4.507	96	7923	4.962	ug/l	89
43) Isopropyl Acetate	6.348	43	21116	5.286	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091321\
Data File : VX024192.D
Acq On : 13 Sep 2021 18:22
Operator : JC/MD
Sample : M3709-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
250EM-DISPOSAL-1

Quant Time: Sep 14 05:25:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091321W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 13 12:27:19 2021
Response via : Initial Calibration

