

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
 Data File : VX037689.D
 Acq On : 13 Sep 2023 12:51
 Operator : JC/MD
 Sample : 04373-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1043-MW-03A(17)DL

Quant Time: Sep 14 02:18:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	231544	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	406052	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	374905	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	160981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	162142	43.282	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.560%
35) Dibromofluoromethane	5.385	113	129600	45.029	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.060%
50) Toluene-d8	8.647	98	492510	45.373	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.740%#
62) 4-Bromofluorobenzene	11.079	95	169426	43.264	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.520%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	8158	5.386	ug/l #	88
27) cis-1,2-Dichloroethene	4.495	96	23710	7.094	ug/l	82
44) Trichloroethene	7.129	130	36610	12.345	ug/l	97
64) Tetrachloroethene	9.275	164	81816	33.683	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
Data File : VX037689.D
Acq On : 13 Sep 2023 12:51
Operator : JC/MD
Sample : 04373-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1043-MW-03A(17)DL

Quant Time: Sep 14 02:18:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

