

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
 Data File : VX037691.D
 Acq On : 13 Sep 2023 13:40
 Operator : JC/MD
 Sample : 04394-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB9/8

Quant Time: Sep 14 02:19:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	225757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	399748	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	369632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	154290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	165759	45.381	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.760%
35) Dibromofluoromethane	5.391	113	130903	46.199	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.400%
50) Toluene-d8	8.647	98	484710	45.359	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.720%#
62) 4-Bromofluorobenzene	11.079	95	164047	42.551	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.100%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3622	2.453	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
Data File : VX037691.D
Acq On : 13 Sep 2023 13:40
Operator : JC/MD
Sample : 04394-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB9/8

Quant Time: Sep 14 02:19:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

