

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
 Data File : VX037692.D
 Acq On : 13 Sep 2023 14:04
 Operator : JC/MD
 Sample : 04373-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1046-FB090623

Quant Time: Sep 14 02:19:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	227637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	391228	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	349504	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	149380	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156128	42.391	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.780%
35) Dibromofluoromethane	5.385	113	124848	45.022	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.040%
50) Toluene-d8	8.653	98	472418	45.171	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.340%#
62) 4-Bromofluorobenzene	11.079	95	157212	41.666	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.340%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	46149	30.994	ug/l	92
25) 2-Butanone	4.574	43	25524	10.513	ug/l	92
51) 4-Methyl-2-Pentanone	8.580	43	10253	2.267	ug/l	97
95) Naphthalene	13.774	128	244401	22.251	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
Data File : VX037692.D
Acq On : 13 Sep 2023 14:04
Operator : JC/MD
Sample : 04373-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1046-FB090623

Quant Time: Sep 14 02:19:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

