

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
 Data File : VX037695.D
 Acq On : 13 Sep 2023 15:18
 Operator : JC/MD
 Sample : 04353-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-090723

Quant Time: Sep 14 02:21:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	231999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	403691	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	365732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	167440	44.608	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.220%
35) Dibromofluoromethane	5.385	113	132572	46.331	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.660%
50) Toluene-d8	8.647	98	494064	45.782	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.560%#
62) 4-Bromofluorobenzene	11.079	95	160227	41.154	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.300%#
Target Compounds						
16) Acetone	2.380	43	6731	4.436	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
Data File : VX037695.D
Acq On : 13 Sep 2023 15:18
Operator : JC/MD
Sample : 04353-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-090723

Quant Time: Sep 14 02:21:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

