

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
 Data File : VX037696.D
 Acq On : 13 Sep 2023 15:42
 Operator : JC/MD
 Sample : 04327-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BR-01-170-090623

Quant Time: Sep 14 02:21:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	217141	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	378722	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	349249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	157863	44.934	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.860%
35) Dibromofluoromethane	5.391	113	123524	46.015	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.040%
50) Toluene-d8	8.647	98	463665	45.798	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.600%#
62) 4-Bromofluorobenzene	11.079	95	151200	41.396	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.800%#
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.313	96	56026	25.079	ug/l	86
24) 1,1-Dichloroethane	3.605	63	10445	2.102	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	50637	16.156	ug/l	83
42) 1,2-Dichloroethane	6.092	62	5745	1.328	ug/l	91
44) Trichloroethene	7.129	130	19430	7.025	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
Data File : VX037696.D
Acq On : 13 Sep 2023 15:42
Operator : JC/MD
Sample : 04327-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BR-01-170-090623

Quant Time: Sep 14 02:21:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

