

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
 Data File : VX037697.D
 Acq On : 13 Sep 2023 16:06
 Operator : JC/MD
 Sample : 04352-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWBR-01-170-090723

Quant Time: Sep 14 02:21:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	241560	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	416080	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	375542	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153244	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	163864	41.928	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.860%
35) Dibromofluoromethane	5.385	113	131451	44.571	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.140%
50) Toluene-d8	8.653	98	505507	45.448	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.900%#
62) 4-Bromofluorobenzene	11.079	95	164814	41.072	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.140%#
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	4045	2.560	ug/l	94
40) Benzene	6.050	78	4704	0.387	ug/l	99
44) Trichloroethene	7.129	130	3045	1.002	ug/l	78
52) Toluene	8.720	92	2665	0.341	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
Data File : VX037697.D
Acq On : 13 Sep 2023 16:06
Operator : JC/MD
Sample : 04352-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWBR-01-170-090723

Quant Time: Sep 14 02:21:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

