

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
 Data File : VX037699.D
 Acq On : 13 Sep 2023 16:55
 Operator : JC/MD
 Sample : 04353-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-12A-15.5-090723

Quant Time: Sep 14 02:22:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	230228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	398792	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	368133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151604	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158930	42.667	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.340%
35) Dibromofluoromethane	5.391	113	128246	45.370	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.740%
50) Toluene-d8	8.653	98	486046	45.593	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.180%#
62) 4-Bromofluorobenzene	11.079	95	161208	41.915	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.820%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.501	96	5870	1.766	ug/l	85
44) Trichloroethene	7.129	130	30186	10.364	ug/l	99
64) Tetrachloroethene	9.275	164	678	0.284	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
Data File : VX037699.D
Acq On : 13 Sep 2023 16:55
Operator : JC/MD
Sample : 04353-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12A-15.5-090723

Quant Time: Sep 14 02:22:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

