

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
 Data File : VX037702.D
 Acq On : 13 Sep 2023 18:09
 Operator : JC/MD
 Sample : 04356-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DISCHARGE-BAY-STREET-SPRING-STREET

Quant Time: Sep 14 02:23:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	221272	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	387931	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	361253	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	148469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	159754	44.624	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.240%
35) Dibromofluoromethane	5.385	113	127348	46.313	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.620%
50) Toluene-d8	8.647	98	480384	46.323	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.640%
62) 4-Bromofluorobenzene	11.079	95	160960	43.022	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.040%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3031	2.094	ug/l #	87
30) Chloroform	5.099	83	3320	0.607	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
Data File : VX037702.D
Acq On : 13 Sep 2023 18:09
Operator : JC/MD
Sample : 04356-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DISCHARGE-BAY-STREET-SPRING-STREET

Quant Time: Sep 14 02:23:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

