

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
 Data File : VX037704.D
 Acq On : 13 Sep 2023 18:57
 Operator : JC/MD
 Sample : PB155499TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB155499TB

Quant Time: Sep 14 02:24:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	217948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	387623	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	373115	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	154274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158251	44.878	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.760%
35) Dibromofluoromethane	5.385	113	125480	45.670	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.340%
50) Toluene-d8	8.647	98	482627	46.576	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.160%
62) 4-Bromofluorobenzene	11.079	95	166169	44.450	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.900%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	103940	72.909	ug/l	94
20) Methylene Chloride	2.788	84	12213	3.239	ug/l	90
25) 2-Butanone	4.568	43	25588	11.008	ug/l	92
43) Isopropyl Acetate	6.342	43	145686	21.353	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
Data File : VX037704.D
Acq On : 13 Sep 2023 18:57
Operator : JC/MD
Sample : PB155499TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB155499TB

Quant Time: Sep 14 02:24:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

