

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
 Data File : VX037713.D
 Acq On : 14 Sep 2023 06:11
 Operator : JC/MD
 Sample : 04399-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-11930

Quant Time: Sep 14 07:28:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	252822	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	443489	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	430071	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	190032	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	177392	43.367	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.740%
35) Dibromofluoromethane	5.379	113	142438	45.312	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.620%
50) Toluene-d8	8.647	98	551009	46.477	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.960%
62) 4-Bromofluorobenzene	11.079	95	196365	45.910	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.820%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	24395	14.752	ug/l #	84
20) Methylene Chloride	2.788	84	14686	3.406	ug/l	88
25) 2-Butanone	4.562	43	14076	5.220	ug/l	94
43) Isopropyl Acetate	6.336	43	147513	18.897	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
Data File : VX037713.D
Acq On : 14 Sep 2023 06:11
Operator : JC/MD
Sample : 04399-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-11930

Quant Time: Sep 14 07:28:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

