

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
 Data File : VX037717.D
 Acq On : 14 Sep 2023 07:48
 Operator : JC/MD
 Sample : 04401-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH110

Quant Time: Sep 14 08:46:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	260115	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	458183	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	433485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	189823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	181708	43.177	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.360%
35) Dibromofluoromethane	5.385	113	148322	45.671	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.340%
50) Toluene-d8	8.647	98	562856	45.954	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.900%#
62) 4-Bromofluorobenzene	11.079	95	196770	44.530	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.060%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	24900	14.635	ug/l	91
20) Methylene Chloride	2.794	84	15093	3.401	ug/l	85
25) 2-Butanone	4.568	43	14161	5.104	ug/l	92
43) Isopropyl Acetate	6.342	43	158745	19.684	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091323\
Data File : VX037717.D
Acq On : 14 Sep 2023 07:48
Operator : JC/MD
Sample : 04401-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH110

Quant Time: Sep 14 08:46:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

