

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091418\
 Data File : VX004531.D
 Acq On : 14 Sep 2018 11:12
 Operator : JC/MD
 Sample : VX0914MBL01
 Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0914MBL01

Quant Time: Sep 15 04:52:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	256812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	359466	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	326551	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	176808	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	174151	52.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.18%	
35) Dibromofluoromethane	5.50	113	156937	47.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.42%	
50) Toluene-d8	8.71	98	496253	44.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.98%	
62) 4-Bromofluorobenzene	11.14	95	167605	42.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.12%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091418\
Data File : VX004531.D
Acq On : 14 Sep 2018 11:12
Operator : JC/MD
Sample : VX0914MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0914MBL01

Quant Time: Sep 15 04:52:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

