

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091418\
 Data File : VX004544.D
 Acq On : 14 Sep 2018 17:07
 Operator : JC/MD
 Sample : J4897-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WASTE-CLASS-DRUMS

Quant Time: Sep 17 15:08:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277024	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	391981	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	367591	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	213749	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	178876	50.08	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.16%	
35) Dibromofluoromethane	5.51	113	162694	45.36	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	90.72%	
50) Toluene-d8	8.71	98	573654	47.69	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.38%	
62) 4-Bromofluorobenzene	11.14	95	202808	47.23	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	94.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	12534	7.19	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091418\
Data File : VX004544.D
Acq On : 14 Sep 2018 17:07
Operator : JC/MD
Sample : J4897-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WASTE-CLASS-DRUMS

Quant Time: Sep 17 15:08:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

