

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091418\
 Data File : VX004547.D
 Acq On : 14 Sep 2018 18:27
 Operator : JC/MD
 Sample : J4913-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPSI-TT-MW303I2-20180912

Quant Time: Sep 17 15:57:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	282818	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	396859	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	356076	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	193652	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	169677	46.53	ug/l	0.00
Spiked Amount			Recovery	=	93.06%	
35) Dibromofluoromethane	5.51	113	159622	43.95	ug/l	0.00
Spiked Amount			Recovery	=	87.90%	
50) Toluene-d8	8.72	98	549101	45.09	ug/l	0.00
Spiked Amount			Recovery	=	90.18%	
62) 4-Bromofluorobenzene	11.14	95	188790	43.43	ug/l	0.00
Spiked Amount			Recovery	=	86.86%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Trichloroethene	7.21	130	5547	1.49	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091418\
Data File : VX004547.D
Acq On : 14 Sep 2018 18:27
Operator : JC/MD
Sample : J4913-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPSI-TT-MW303I2-20180912

Quant Time: Sep 17 15:57:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

