

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091418\
 Data File : VX004565.D
 Acq On : 15 Sep 2018 05:49
 Operator : JC/MD
 Sample : J4925-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S13-0255

Quant Time: Sep 17 01:02:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	264201	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	365704	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	341433	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	183848	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	174582	51.25	ug/l	0.00
Spiked Amount			Recovery	=	102.50%	
35) Dibromofluoromethane	5.50	113	158978	47.51	ug/l	0.00
Spiked Amount			Recovery	=	95.02%	
50) Toluene-d8	8.71	98	514602	45.85	ug/l	0.00
Spiked Amount			Recovery	=	91.70%	
62) 4-Bromofluorobenzene	11.14	95	177502	44.31	ug/l	0.00
Spiked Amount			Recovery	=	88.62%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	7852	4.72	ug/l	95
31) Cyclohexane	5.58	56	1263	0.26	ug/l	93
39) Methylcyclohexane	7.46	83	1982	0.43	ug/l #	80
67) Ethyl Benzene	10.25	91	7258	0.53	ug/l	97
73) Isopropylbenzene	11.02	105	3258	0.27	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	3963	0.38	ug/l	97
85) sec-Butylbenzene	11.94	105	3423	0.28	ug/l	85
95) Naphthalene	13.83	128	7520	0.60	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091418\
Data File : VX004565.D
Acq On : 15 Sep 2018 05:49
Operator : JC/MD
Sample : J4925-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S13-0255

Quant Time: Sep 17 01:02:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

