


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091418\
Data File : VX004553.D
Acq On : 14 Sep 2018 23:55
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Title : SW846 8260
Last Update : Wed Sep 12 02:17:23 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	2520	PASS
75	95	30	60	46.3	6839	PASS
95	95	100	100	100.0	14761	PASS
96	95	5	9	6.4	946	PASS
173	174	0.00	2	1.2	177	PASS
174	95	50	100	99.3	14652	PASS
175	174	5	9	7.8	1144	PASS
176	174	95	101	95.2	13956	PASS
177	176	5	9	6.4	888	PASS