

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091420\
 Data File : VX018388.D
 Acq On : 14 Sep 2020 16:58
 Operator : JC/SP
 Sample : PB131574ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131574ZHE#04

Quant Time: Sep 15 05:55:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	448575	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	699473	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	717222	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	351984	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	314789	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
35) Dibromofluoromethane	5.47	113	246402	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
50) Toluene-d8	8.70	98	878466	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
62) 4-Bromofluorobenzene	11.12	95	343634	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
Target Compounds						
16) Acetone	2.42	43	27696	11.414	ug/l	92
18) Methyl Acetate	2.75	43	6989	1.194	ug/l #	73
20) Methylene Chloride	2.83	84	14328	1.761	ug/l	90
43) Isopropyl Acetate	6.42	43	123461	10.201	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091420\
Data File : VX018388.D
Acq On : 14 Sep 2020 16:58
Operator : JC/SP
Sample : PB131574ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131574ZHE#04

Quant Time: Sep 15 05:55:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

