

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091420\
 Data File : VX018393.D
 Acq On : 14 Sep 2020 18:52
 Operator : JC/SP
 Sample : PB131574TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131574TB

Quant Time: Sep 15 06:01:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	445401	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	709987	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	713721	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	358380	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	325238	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
35) Dibromofluoromethane	5.47	113	243583	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
50) Toluene-d8	8.70	98	904252	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
62) 4-Bromofluorobenzene	11.12	95	345509	48.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.52%	
Target Compounds						
16) Acetone	2.42	43	28885	11.988	ug/l #	87
18) Methyl Acetate	2.75	43	8169	1.405	ug/l	99
20) Methylene Chloride	2.84	84	12930	1.492	ug/l	92
43) Isopropyl Acetate	6.42	43	122671	9.986	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091420\
Data File : VX018393.D
Acq On : 14 Sep 2020 18:52
Operator : JC/SP
Sample : PB131574TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131574TB

Quant Time: Sep 15 06:01:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

