

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091420\
 Data File : VX018395.D
 Acq On : 14 Sep 2020 19:38
 Operator : JC/SP
 Sample : L3863-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HD-02-091120

Quant Time: Sep 15 06:05:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	459589	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	748120	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	717604	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	359026	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	329466	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
35) Dibromofluoromethane	5.46	113	253779	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
50) Toluene-d8	8.70	98	927320	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
62) 4-Bromofluorobenzene	11.12	95	346448	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
Target Compounds						
16) Acetone	2.42	43	23782	9.566	ug/l	90
18) Methyl Acetate	2.75	43	7480	1.247	ug/l	91
20) Methylene Chloride	2.84	84	13438	1.511	ug/l	93
43) Isopropyl Acetate	6.41	43	122277	9.446	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091420\
Data File : VX018395.D
Acq On : 14 Sep 2020 19:38
Operator : JC/SP
Sample : L3863-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HD-02-091120

Quant Time: Sep 15 06:05:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

