

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091420\
 Data File : VX018398.D
 Acq On : 14 Sep 2020 20:46
 Operator : JC/SP
 Sample : L3993-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAV-091120-DP

Quant Time: Sep 15 06:18:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	432567	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	680833	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	685321	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	334438	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	308477	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
35) Dibromofluoromethane	5.47	113	244434	52.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.44%	
50) Toluene-d8	8.70	98	877644	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	11.12	95	325193	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
Target Compounds						
16) Acetone	2.42	43	19174	8.194	ug/l #	88
18) Methyl Acetate	2.75	43	8348	1.478	ug/l	98
20) Methylene Chloride	2.84	84	12561	1.492	ug/l	91
43) Isopropyl Acetate	6.42	43	128725	10.927	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091420\
Data File : VX018398.D
Acq On : 14 Sep 2020 20:46
Operator : JC/SP
Sample : L3993-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAV-091120-DP

Quant Time: Sep 15 06:18:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

