


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091420\
Data File : VX018374.D
Acq On : 14 Sep 2020 09:57
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Title : SW846 8260
Last Update : Fri Aug 21 03:03:56 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1635

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	24130	PASS
75	95	30	60	51.3	71442	PASS
95	95	100	100	100.0	139272	PASS
96	95	5	9	6.0	8339	PASS
173	174	0.00	2	0.5	609	PASS
174	95	50	100	80.7	112346	PASS
175	174	5	9	7.6	8548	PASS
176	174	95	101	97.0	108994	PASS
177	176	5	9	5.8	6273	PASS