

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024209.D
 Acq On : 14 Sep 2021 15:54
 Operator : JC/MD
 Sample : M3729-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-TB-20210908

Quant Time: Sep 15 01:57:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	144489	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	243967	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100357	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103883	50.420	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.840%	
35) Dibromofluoromethane	5.397	113	82197	50.238	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.480%	
50) Toluene-d8	8.653	98	300489	50.414	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.820%	
62) 4-Bromofluorobenzene	11.085	95	111602	48.113	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.220%	
Target Compounds						
						Qvalue
20) Methylene Chloride	2.794	84	1007	0.597	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024209.D
Acq On : 14 Sep 2021 15:54
Operator : JC/MD
Sample : M3729-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-TB-20210908

Quant Time: Sep 15 01:57:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

